



Product Service

# Attestation of Compliance

No. E8N 09 08 62545 005

**Holder of Certificate:** **DA TECH CO., Ltd**  
 #710-C, Woolim Lionsvalley, 425  
 Cheongcheon-dong, Bupyeong-gu, Incheon 403-858  
 REPUBLIC OF KOREA

**Name of Object:** Electrical equ. for measurement, control and laboratory use  
 Digital Breathalyzer Alcoscent

**Model(s):** DA-8100

**Description of Object:** Rated input voltage: 3 V d.c. (1.5 V Battery \* 2EA)

**Tested according to:** EN 61326-2-3:2006 Class B

This Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

**Test report no.:** ESTCE-0908-003

**Date,** 2009-08-13



**CE** After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.